Ricardo Morren, Operations Manager
Juvo Solutions
18 Mallaig Way,
Canningvale, WA 6155

7th July 2015

Dear Ricardo,

Reference: testing the Juvo Solutions JH1071attendant propelled mobile shower commode (710 mm wide) to 400 kg, our Laboratory reference is 492466.

This letter is to advise you that this Laboratory has finished testing your product to the following Standard method: AS/NZS 3973:2009 Shower / toilet chairs – Mobile and static (except corrosion resistance and surface finishes).

The official ten (10) page report, dated 6th July 2015 and referencing job number 492466 identifies the sample, the test documents and the testing facility. All conformances with the AS/NZS standards are indicated in the report.

The Juvo Solutions JH1071 attendant propelled mobile shower commode (710 mm wide) as represented by the test sample, including the supplied documentation and labelling as stated in the test report, complies with the relevant requirements of AS/NZS 3973:2009 Shower / toilet chairs – Mobile and static (excluding the methods indicated in the report as “not tested”) for user mass 400 kg.

Compliance is conditional that all products made available for sale must be marked with the maximum user weight (400 kg)

Please do not hesitate to contact us if you have any enquiries or comments regarding the test.

Kind regards,

Digitally signed by Wayne Wurfel
DN: cn=Wayne Wurfel, o=Test Laboratory, ou=NovitaTech, email=wayne.wurfel@novita.org.au, c=AU
Reason: I attest to the accuracy and integrity of this document
Date: 2015.07.08 08:23:20 +09'30'

Wayne Wurfel
Senior Test Technician,
Authorised signatory

NovitaTech (a division of Novita Children’s Services)
171 Days Road, Regency Park, South Australia 5010
PO Box 2438, Regency Park, South Australia 5942
Telephone 1300 855 585 • Fax (08) 8243 8208
enquiries@novita.org.au • www.novitatech.org.au
ABN 53 137 537 636

Quality System
Quality Endorsed Company
ISO 9001 QEC8382